



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: 4/24/2006 INITIALS: Dec OWNER: WSP
DISTRICT: 1 ADDRESS: 7th St
LOCATION: Shop COUNTY:
MAKE: Hamm TYPE: port wheel weigher ACCURACY CLASS: III
MODEL: 101 ID. NO: 23139 CAPACITY: 20000 SECTIONS: 1
MINIMUM GRADUATIONS: 50 REG. ELEMENT: dial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10050	+50				
"	15000	15050	+50				
Center	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: 4/26/2006


State of Washington
County of Thurston

Signed and sworn to (or affirmed) before
me on 4-26-06 by

Michael J. Evans
(Signature of notary public)

Michael J. EVANS

Printed Name of Notary
My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS